



EU Declaration of Conformity

Product	Laboratory Equipment	File Number	P225
Manufacturer	Cole-Parmer Ltd Beacon Road Stone, Staffordshire ST15 0SA United Kingdom		

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration Gelation Timer
(reference the attached list of catalogue numbers)

The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation:

Low Voltage Directive	2006/95/EC	(until 19 April 2016)
	2014/35/EU	(from 20 April 2016)
EMC Directive	2004/108/EC	
RoHS Directive	2011/65/EC	

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

IEC/EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control and laboratory use. Part 1: General requirements.
IEC/EN 61010-2-051:2015	Particular requirements for laboratory equipment for mixing and stirring.
IEC/EN 61326-1:2006	Electrical equipment for measurement, control and laboratory use. EMC requirements. Part 1: General requirements (Class B).

Signed for and on behalf of the above manufacturer

Additional Information Year of CE Marking: 1996

Place of Issue Stone, Staffordshire, UK

Date of Issue June 2009, Revised 4 September 2017

Authorised Representative Carl Warren

Title Technical Manager

Signature

Range	Catalogue Number	Part Code	Description
GELTM	98930-40	FGT6/120V/60	Gelation Timer, 999.9 minutes, 120V, 60Hz
GELTM	98930-41	FGT6	Gelation Timer, 999.9 minutes, 230V, 50Hz
GELTM	98930-42	FGT5/120V/60	Gelation Timer, 9999 minutes, 120V, 60Hz
GELTM	98930-43	FGT5	Gelation Timer, 9999 minutes, 230V, 50Hz